

# **AZIDE DEXTROSE BROTH, ROTHE**

A selective medium for the detection of enterococci.

Dehydrated media	
Code number:	500 g: ADR20500, 5 kg: ADR25000
Colour:	Yellowish
Appearance:	Homogeneous hygroscopic powder
pH before autoclaving (25 °C):	6,8 - 7,2

**Direction:** Suspend **36 g** in one litre of distilled water and heat gently to dissolve the medium completely. Dispense into final containers and sterilise by autoclaving at 115 °C for 15 minutes.

### Warning!

The medium is heat sensitive. No further sterilisation is necessary or desirable.

Prepared media	
Bottled media:	100 ml: ADR30100, 500 ml: ADR30500
Tubed media:	150 x 15 mm: ADR40010 (10 ml)
Colour:	Yellowish
pH (25 °C):	6,9 - 7,1

**Direction:** Dispense the bottled media aseptically into sterile final containers. Media in tubes are ready to use.

## FORMULA in g/l

Peptones	20,4
Glucose	5,0
Sodium chloride	5,0
Sodium azide	0,2
Buffers	5,4

**Note:** The typical formula can be adjusted to obtain optimal performance.

**Storage conditions:** Store the dehydrated media tightly closed in a dry place at room temperature. Store the bottled and tubed media protected from light at room temperature. Use before the expiry date on the label.

#### Quality control:

Test strains	Incubation temp: 37 °C	Growth	Incubation time: 24 h
Enterococcus faecalis	ATCC 29212	Good	
Escherichia coli	ATCC 25922	Inhibited	

**References:** Greenberg et al. (1985) APHA, Standard Methods for the Examination of Water and Wastewater, 16<sup>th</sup> ed.

### In vitro diagnostic - for professional use only!