

ENDO AGAR, DEV

A selective and differential medium for the detection of coliforms and other enteric bacteria according to DEV.

Dehydrated media	
Code number:	500 g: EDE20500, 5 kg: EDE24600
Packaging of 500 g:	500 g medium base + 1 x 45 ml fuchsin solution
Packaging of 5 kg:	5 kg medium base + 4x100 ml fuchsin solution
Appearance of agar base:	Yellowish, homogeneous hygroscopic powder
Appearance of supplement:	Deep magenta, precipitation free solution
pH before autoclaving (25 °C):	6,7 - 7,1

Direction: Suspend **58 g** in one litre of distilled water. Add **5 ml of Endo Basic Fuchsin Solution.** Mix well and heat with frequent agitation until the medium boils well. Sterilise by autoclaving at 121 °C for 15 minutes. Mix well again before pouring.

Prepared media		
Bottled media:	100 ml: EDE30100, 500 ml: EDE30500	
Plated media:	55 mm: EDE50055, 90 mm: EDE50090	
Colour:	Pale pink	
pH (25 °C):	6,8 - 7,0	

Direction: Dispense the melted bottled media aseptically into sterile Petri-dishes. Media in Petri-dishes are ready to use.

FORMULA FOR COMPLETE MEDIUM in g/l

Peptones	20,5
Lactose	10,0
Sodium chloride	5,0
Sodium sulphite	2,5
Basic fuchsin	0,5
Agar	20,0

Note: The typical formula can be adjusted to obtain optimal performance.

Storage conditions: Store the dehydrated media tightly closed in a dry place at room temperature. Store the bottled media protected from light at room temperature. Store the supplements and the plated media protected from light at 2-8 °C. Use before the expiry date on the label.

Quality control:

Test strains	Incubation temp: 37 °C	Growth	Incubation time: 24 h
Escherichia coli ATCC 25922		Good, dark red colonies with green metallic sheen	
Salmonella typhimurium ATCC 14028		Good, colourless colonies	
Enterococcus faecalis ATCC 29212		Inhibited	

References: DIN 38411

In vitro diagnostic - for professional use only!